

# ARCS Instrument Replacement Form

System: Anemomter WND1 (High)\_\_\_ **Replacement**

Fill out the following form when replacing instruments at an ARCS site:

Site: Nauru\_\_\_\_\_

## OLD Configuration

Part Number	Serial Number	Location	Configuration Version	Change Item 14	Change Item 15	Date GMT	Time Start GMT	Time End GMT
05106/08234	184825/44678	SMET2	V981028.00	0.09758	0.33	03/13-14/99		

## NEW Configuration

Part Number	Serial Number	Location	Configuration Version	Change Item 14	Change Item 15	Copy Config. Logger to PC	Save to EPROM(E)	Update Version #
05106/08234	31103/58318	SMET2	V990314.00	0.09642	0.38	X	X	X

Replacement completed by: Fred Helsel\_\_\_\_\_

## **NOTES:**

Old Part remains on site as spare: ☐

Old Part packed for shipment to: AIS\_\_\_\_\_

Other old part disposition (explain):

Other NOTES:

Replace anemometer for repair (speed dropouts observed)